

TEST REPORT

TUCHENG

Date of Issue: 27 Dec 2023

TFG2L197

Report No.:

Date of Receipt and Test Start:

18 Dec 2023

Page Order/Pages:

Report Title : AceGreen Eco-Material Technology Co., Ltd.(U0940) Address : No. 50, Ln. 20, Sec. 1, Nantong Rd., Ershui Township, Changhua County 530, Taiwan Item : Mask Quantity: 1 PC Ref. No. : --Sample description is given by the client :

Sample 1 : AceGreen Medical Face Mask

Test Results :

T. (I.	Test Results	Test Method
Test Items	Sample1	
Free Formaldehyde Content	N.D.	CNS 14774 T5017-
(ppm)		2022 9.7/CNS 15580-1
	Resear	L3271-1-2012

Note: N.D.=Not Detected.(<16ppm).

Note: Limit value: \leq 75 ppm

Test Items	Test Results	Test Method
	Sample1	
Test for Migratable Fluorescent Substances	No	CNS 14774 T5017-
		2022 9.9/CNS 11820
		P3082-2007

Authorized by president of Taiwan Textile Research Institute

Director, Department of Testing and Certification



Note:

1. This report is only responsible for the submitted sample(s), which will be kept for one month period. 2. This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3. The test report should not be used for public advertisement and commercial promotion.

Department of Testing and Certification, Taiwan Textile Research Institute No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R. O. C.) Tel : +886-2-22670321 ext. 7107, 7110 Fax : +886-2-22675108 , +886-2-22689839





1/3



TEST R	REPORT
--------	--------



Report No.: TFG2L197

Date of Issue: 27 Dec 2023

Date of Receipt and Test Start:

t: 18 Dec 2023

Page Order/Pages:

2/3

		Test F	Results	Test Method
	Test Items	Sample1-Pink		
	ľ	Method T	Method D	-
*AZO DYES	4-AMINODIPHENYL	N.D.	N.D.	CNS 14774 T5017-
(mg/kg)	BENZIDINE	N.D.	N.D.	2022 9.8/CNS 1611
	4-CHLORO-O-TOLUIDINE	N.D.	N.D.	1 L3280-1:2019
	2-NAPHTHYLAMINE	N.D.	N.D.	9.1(Method D) 、
	o-AMINOAZOTOLUENE	N.D.	N.D.	9.2(Method T) GC-
	2-AMINO-4-NITROTOLUENE	N.D.	N.D.	MS
	4-CHLOROANILINE	N.D.	N.D.	
2,4-DIAMINOANISOLE 4,4'-	2,4-DIAMINOANISOLE	N.D.	N.D.	
	4,4'-	N.D.	N.D.	
	DIAMINODIPHENYLMETHANE		1.58	
3,3'-DICHLOROBENZIDINE 3,3'-DIMETHOXYBENZIDINE 3,3'-DIMETHYLBENZIDINE 3,3'-DIMETHYL-4,4'-DIAMINC DIPHENYLMETHANE P-CRESIDINE 4,4'-METHYLENE-BIS-(2- CHLOROANILINE) 4,4'-OXYDIANILINE 4,4'-OXYDIANILINE 4,4'-THIODIANILINE 0-TOLUIDINE 2,4-TOLUYLENEDIAMINE 2,4,5-TRIMETHYLANILINE 0-ANISIDINE P-AMINOAZOBENZENE 2,4-Xylidine 2,6-Xylidine	3,3'-DICHLOROBENZIDINE	N.D.	N.D.	
	3,3'-DIMETHOXYBENZIDINE	N.D.	N.D.	
	3,3'-DIMETHYLBENZIDINE	N.D.	N.D.	
	3,3'-DIMETHYL-4,4'-DIAMINO-	N.D.	N.D.	
	DIPHENYLMETHANE		2	A
	P-CRESIDINE	N.D.	N.D.	
	4,4'-METHYLENE-BIS-(2-	N.D.	N.D.	
	CHLOROANILINE)	FDI.		
	4,4'-OXYDIANILINE	N.D.	N.D.	
	4,4'-THIODIANILINE	N.D.	N.D.	
	O-TOLUIDINE	N.D.	N.D.	
	2,4-TOLUYLENEDIAMINE	N.D.	N.D.	
	2,4,5-TRIMETHYLANILINE	N.D.	N.D.	
	O-ANISIDINE	N.D.	N.D.	
	P-AMINOAZOBENZENE	N.D.	N.D.	
	2,4-Xylidine	N.D.	N.D.	
	2,6-Xylidine	N.D.	N.D.	

Note: 1. "*"Subcontracted test. 2. N.D.=Not Detected.(<5 mg/kg) 3. Limit of Standard





TEST REPORT T



Report No.: TFG2L197

Date of Issue: 27 Dec 2023

- Date of Receipt and Test Start:
- 18 Dec 2023

Page Order/Pages:

3/3

Sample photo :

